



Preventative Maintenance Schedule

All maintenance of the Life Pulse High-Frequency Ventilator must be performed by Bunnell trained service personnel or by a hospital technician that has attended the Bunnell Service Seminar.

Bunnell Incorporated recommends that the Life Pulse and Patient Box receive preventive maintenance every 12 months or 2,000 hours of operation, whichever comes first.

As with any calibrated equipment, consistent and accurate performance is dependent on regularly scheduled maintenance.

Service Contracts for preventive maintenance and service beyond the standard warranty periods are available through Bunnell Incorporated.

For more information contact our Customer Service Department during normal business hours (8:00 a.m. to 5:00 p.m. Mountain Time) at 800.800.4358 ext. 6.